

LS3300 AC power calibrator

Technical Information



Contents

- 1. Introduction**
- 2. Basic Characteristic Measurement Data**
 - 2-1. Stability**
 - 2-2. Frequency characteristic**
 - 2-3. Linearity**
 - 2-4. Distortion rate**
 - 2-5. Output noise**
 - 2-6. Output response**
 - 2-7. Compliance**
 - 2-8. Communication speed (GPIB, Ethernet and USB)**
- 3. Synchronous Operation**
 - 3-1. Power error at three-phase power calibration**
 - 3-2. Stability of large 1P2W current**
 - 3-3. Example of calibration using AUX terminal**
 - 3-4. Change of wiring system via communication and setting procedure**
- 4. Others**
 - 4-1. Safety of output terminal**
 - 4-2. Example of error cause during voltage/current generation**
 - 4-3. Maximum capacity and inductive load**
 - 4-4. Power meter in which voltage and current terminals are connected internally**
 - 4-5. Example of calibration of clamp-on power meter**
 - 4-6. Indication of output value**
 - 4-7. Sweep behavior**
 - 4-8. LINE synchronization**

1. Introduction

This document mainly provides the characteristic data about items that are not specified in the specifications of the LS3300.

From the nature of the data, please consider the measurement results described in this document to be reference values.

For the specifications guaranteed by the LS3300, please refer to the specifications described in the catalog and the user's manual.

The data included in this document is measured at the reference temperature ($23\text{ }^{\circ}\text{C} \pm 3\text{ }^{\circ}\text{C}$) unless otherwise noted. The measurement instruments used for creation of this document are as follows:

| Measurement instrument | Model | Manufacturer |
|---------------------------|---------|--------------|
| AC Power Calibrator | LS3300 | YOKOGAWA |
| Precision Power Analyzer | WT3000E | YOKOGAWA |
| AC/DC Current Sensor | CT200 | YOKOGAWA |
| AC/DC Current Sensor | CT1000 | YOKOGAWA |
| Mixed Signal Oscilloscope | DLM2054 | YOKOGAWA |
| Universal Counter | TC110 | YOKOGAWA |
| Reference Multi Meter | 8508A | FLUKE |
| Phase Meter | 6000A | Clarke-Hess |

The current transformation ratio of the CT200 and the CT1000 is 1000:1 and 1500:1, respectively.

The LS3300 is set as follows unless otherwise noted:

Wiring = 1P2W

Oscillation = INT

Lo to Earth = ON

2-1-1. Stability of voltage and current

■ Overview

This is the fluctuation of the output error of a representative range when the LS3300 is connected to the power meter WT3000E.

The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Power Meter WT3000E

■ Device settings


WT3000E Update rate = 500 ms

LS3300 Range = 100 V, 10 A, AUX 5 V

Output setting value = 100% of each range

Frequency = 60 Hz

■ Measurement result



2-1-2. Stability of large current output

■ Overview

This is the fluctuation of the output error of the 50 A range when the LS3300 is connected to the power meter WT3000E through the CT200. The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Power Meter WT3000E

Current Sensor CT200

■ Device settings


WT3000E Update rate = 500 ms

LS3300 Range = 50 A

Output setting value = 100% of each range

Frequency = 60Hz

■ Measurement result



2-1-3. Power stability (1)

■ Overview

This is the fluctuation of the output error at the phase angle setting of power factor(λ)=0 or 1 when the LS3300 is connected to the power meter WT3000E and outputs 500 VA.

The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Power Meter WT3000E

2-1-3. Power stability (2)

■ Device settings

WT3000E Voltage range = 100 V, Current range = 5 A

Update rate = 500 ms



LS3300 Range = 100 V, 10 A

Output setting value = 100 V, 5 A

Frequency = 60 Hz

■ Measurement result

➤ PF= 1



2-1-3. Power stability (3)

■ Device settings

WT3000E Voltage range = 100 V, Current range = 5 A

Update rate = 500 ms


LS3300 Range = 100 V, 10 A

Output setting value = 100 V, 5A

Frequency = 60 Hz

■ Measurement result

➤ PF= 0



2-1-4. Frequency stability

■ Overview

This is the fluctuation of the frequency error of a representative range when the LS3300 is connected to the TC110.

The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Universal Counter TC110

■ Device settings

TC110 CH = A


FILT ON

LS3300 Range = 10 V

Output setting value = 100% of each range

Frequency = 1 kHz

■ Measurement result



2-1-5. Phase stability (1)

■ Overview

This is the fluctuation of the phase at the phase angle setting of power factor(λ)=0 when the LS3300 is connected to the power meter WT3000E and outputs 100 V/10 A and 100 V/AUX 5 V.
The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device



Power Meter WT3000E

■ Device settings

WT3000E Voltage range = 100 V, Current range = 10 A, EXT 5 V
Update rate = 500ms
LS3300 Range = 100 V, 10 A, AUX 5 V
Output setting value = 100 V, 10 A, AUX 5 V
Frequency = 60Hz
Power factor = 0

2-1-5. Phase stability (2)

■ Measurement result



2-1-5. Phase stability (3)

■ Overview

This is the fluctuation of the phase between the devices when two LS3300s are connected to the 6000A and output 100V.

The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Phase Meter Clarke-Hess 6000A

■ Device settings


LS3300

Master & Slave Range = 100 V, Output setting value = 100% setting,
Wiring system = 1P2W

Master Reference signal = INT, Frequency = 60 Hz, Phase = 0°

Slave Reference signal = EXT, Phase = 0°, 30°, 60°, 90°

■ Measurement result




2-2. Frequency characteristic

- Overview
This is the frequency characteristic of LS3300's output accuracy.
The output errors of each frequency are plotted.
- Used device
Power Meter WT3000E
- Device settings
WT3000E Update rate = 500 ms
LS3300 No load (direct connection to the power meter)
 Range = 100 V, 10 A, AUX 5 V
 Output setting value = 100% of each range
 Wiring system = 1P2W
 Reference signal = INT
 Frequency = 40, 50, 60, 100, 200, 400, 600, 800, 1000,
 1200 Hz


2-2-1. Frequency characteristic (Voltage)

■ Measurement result




2-2-2. Frequency characteristic (Current)

■ Measurement result




2-2-3. Frequency characteristic (AUX)

■ Measurement result



2-2-4. Frequency characteristic (Power)

■ Measurement result



2-2-4. Frequency characteristic (Phase)

■ Overview

This is the frequency characteristic of phase error between the devices when two LS3300s are connected to the 6000A and output 100V.
The plotted data is the average of 10 measurements.

■ Used device

Phase Meter Clarke-Hess 6000A

■ Device settings

LS3300

Master & Slave Range = 100 V, Output setting value = 100% setting,
Wiring system = 1P2W

Master Reference signal = INT

Frequency = 40, 50, 60, 100, 200, 400, 600, 800, 1000, 1200 Hz

Phase = 0°

Slave Reference signal = EXT, Phase = 0°


Specs

50/60 Hz : $\pm 0.03^\circ$

40 to 400 Hz : $\pm 0.10^\circ$

400 to 1.2 Hz : $\pm 0.40^\circ$

■ Measurement result



2-3. Linearity (1)

■ Overview

This is the relationship between a specified value and the output error of a representative range.

It also shows the measurement result by the Fluke8508A regarding the voltage range.

For measurement by the WT3000E and the Fluke8508A, an optimum range is used according to the output value of the LS3300.

■ Used device

Power Meter WT3000E

Multi Meter Fluke8508A

■ Device settings

WT3000E Update rate = 500 ms

LS3300 No load (direct connection to the power meter)


Range = 1 V, 100 V, 10 A

Output setting value = 1, 10, 20, 40, 50, 60, 80, 100, 120% of each range

Frequency = 60Hz

■ Measurement result

➤ Measurement by the WT3000E



2-3. Linearity (2)

■ Device settings

LS3300 Range = 1 V, 100 V

Output setting value = 1, 10, 20, 40, 50, 60, 80, 100, 120%
of each range


Frequency = 60 Hz

8508A ACV Filter = 40 Hz

ACV Resolution = 6

■ Measurement result

➤ Measurement by the Fluke8508A



2-4. Distortion rate (1)

■ Overview

This is the LS3300's distortion rate of a representative range (Measurement values by the WT3000E, Uthd and Ithd).

■ Used device

Power Meter WT3000E

■ Device settings

WT3000E Measurement mode = Wide-band harmonic measurement mode

Measurement order = 0 to 100

Arithmetic expression of distortion rate = $1/\text{Fundamental}$


LS3300 No load (direct connection to the power meter)

Range = 100 V, 10 A, AUX 5 V



Output setting value = 100% of each range

Frequency = 40, 60, 100, 200, 400, 600, 800, 1000, 1200 Hz

■ Measurement result



2-4. Distortion rate (2)



2-5. Output noise (1)

■ Overview

This is the waveforms of output noise observed by the oscilloscope DLM2054 when the setting value of a representative range in the LS3300 is set at 0.

■ Used device

Oscilloscope DLM2054

Banana-BNC Cable, Alligator-BNC Cable (50 Ω , 1 m)

■ Device settings

DLM2054 Vertical axis = 2 mV/div

Horizontal axis = 10 ms/div

Probe = 1:1

Coupling = DC 1 M Ω

Bandwidth = 1 MHz


Acquisition Length = 125 k

LS3300 Range = 10 V, 1 A (10 Ω load), AUX 500 mV

Output setting value = 0% of each setting


Frequency = 60 Hz

■ Connection diagram



2-5. Output noise (2)


■ Measurement result



2.00 mV/div
10 ms/div

P-P(C1) 1.7 mV
Rms(C1) 189.2 μ V
Sdev(C1) 172.8 μ V

0 V output (10 V Range)




200 μ A/div
10 ms/div

P-P(C1) 213 μ A
Rms(C1) 22.3 μ A
Sdev(C1) 22.0 μ A

0 A output (1 A Range)
10 Ω load is connected

2-5. Output noise (3)

■ Measurement result




2.00 mV/div
10 ms/div

P-P(C1) 0.7 mV
Rms(C1) 78.8 μ V
Sdev(C1) 78.5 μ V



0 mV output (AUX 500 mV Range)

- Fluctuations in the p-p and rms values of noise according to the change of bandwidth
The following graph is the measurement result of the p-p and rms values of noise changing the bandwidth of the DLM2054.
Bandwidth = 62.5 kHz, 125 kHz, 250 kHz, 500 kHz, 1 MHz, 2 MHz, 5 MHz, 10 MHz



2-5. Output noise (4)

- Fluctuations in the p-p and rms values of noise according to the change of bandwidth



2-6. Output response (1)

■ Overview

This is the output response of the LS3300 observed by the WT and a oscilloscope.

■ Used device

Power Meter WT3000E

Oscilloscope DLM2054

■ Device settings

WT3000E Update rate = 50 ms

DLM2054 Horizontal axis = 200 ms/div

Coupling = DC 1 M Ω

Bandwidth = 1 MHz

Acquisition Length = 125 k


LS3300 Range = 100 V, 10 A, AUX 5 V

Output setting value = 100% of each range


Frequency = 60 Hz

■ Measurement result (WT3000E)

0 to 100% output of 100 V range



Scope of $\pm 0.05\%$ of final value (100 V)




2-6. Output response (2)

■ Measurement result (DLM2054)

100 V range

Output ON


Output OFF



10 A range

Output ON


Output OFF



AUX 5 V range

Output ON

Output OFF



2-7. Compliance (1)

■ Overview

This is the comparison between maximum output values and specification values when an oscilloscope measures the current of the voltage output terminal during voltage output and the voltage of the current output terminal during current output regarding each range of the LS3300.

The maximum output of the 2558A is also shown as the object of a comparison.

■ Used device

Oscilloscope DLM2054

■ Device settings

LS3300 Range = 1 V, 10 V, 30 V, 100 V, 300 V, 1000 V,
30 mA, 100 mA, 1 A, 10 A, 50 A

Output setting value = 100% of each range

Frequency = 60 Hz

2558A Range = 1 V, 10 V, 100 V, 300 V, 1000 V, 100 mA, 1 A, 10 A,
50 A

Output setting value = 100% of each range

Frequency = 60 Hz

2-7. Compliance (2)

■ Measurement result

Reference: YOKOGAWA 2558A

| Range | Setting value | Product spec | Actual value | Product spec | Actual value |
|--------|---------------|----------------|--------------|---------------|--------------|
| 1 V | 1.00000 V | 0.5 A or more | 1.69 A | 0.5 A or more | 3.4 A |
| 10 V | 10.0000 V | Approx. 600 mA | 945 mA | Approx. 3 A | 4.4 A |
| 30 V | 30.0000 V | Approx. 60 mA | 96 mA | - | - |
| 100 V | 100.000 V | Approx. 60mA | 94 mA | Approx. 0.3 A | 0.44 A |
| 300 V | 300.000 V | Approx. 20 mA | 31.5 mA | Approx. 0.1 A | 0.14 A |
| 1000 V | 1000.00 V | Approx. 6 mA | 9 mA | Approx. 6 mA | 10 mA |
| Range | Setting value | Product spec | Actual value | Product spec | Actual value |
| 30mA | 30.0000mA | Approx. 15V | 21.4V | - | - |
| 100 mA | 100.000m A | Approx. 15 V | 15.7 V | ←Equivalent | |
| 1 A | 1.00000 A | Approx. 15 V | 22.1 V | | |
| 10 A | 10.0000 A | Approx. 3 V | 3.8 V | | |
| 50 A | 50.000 A | Approx. 0.6 V | 0.68 V | | |

2-7. Compliance (3)

■ Overview

This is the influence of cable routing in the case of large current and high frequency.

The voltage between current terminals is measured when there are twists on an attached current cable and no twist.

■ Used device

DMM 73402

B8506WA (attached current cable)

■ Device settings

LS3300 Range = 50 A

Output setting value = 100% of the range

Frequency = 1000 Hz

■ Measurement result



Without twist



With twists

| Cable | Voltage between current terminals |
|---------------|-----------------------------------|
| Without twist | 415 mV |
| With twists | 197 mV |

*The spec of the maximum output is approx. 600 mV

2-8-1. Communication speed (GPIB)

■ Overview

The processing speed of communication commands are as follows:

■ Measurement environment

OS: Windows 7 Enterprise SP1 64-bit

Processor: Intel(R) Core™ i5-5300U CPU @ 2.30GHz 2.30GHz

Memory: 16 GB

Tool: Visual Studio 2015 VC++

■ GP-IB

- When changing the range with “:SOUR:VOLT:RANG”

| | | After change | | | | | |
|---------------|-------|--------------|---------|---------|---------|---------|---------|
| | | 1V | 10V | 30V | 100V | 300V | 1000V |
| Before change | 1V | - | 9.50 ms | 9.51 ms | 9.36 ms | 9.21 ms | 9.52 ms |
| | 10V | 9.41 ms | - | 9.46 ms | 9.52 ms | 9.82 ms | 9.30 ms |
| | 30V | 9.29 ms | 9.42 ms | - | 9.33 ms | 9.59 ms | 9.13 ms |
| | 100V | 9.97 ms | 9.16 ms | 9.32 ms | - | 9.31 ms | 9.33 ms |
| | 300V | 9.35 ms | 9.33 ms | 9.33 ms | 9.42 ms | - | 9.37 ms |
| | 1000V | 9.34 ms | 9.35 ms | 9.50 ms | 9.27 ms | 9.39 ms | - |

- When changing the range with “:SOUR:CURRE:RANG”

| | | After change | | | | | | |
|---------------|------------|--------------|----------|----------|----------|----------|------------|----------|
| | | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
| Before change | 30mA | - | 9.72 ms | 9.86 ms | 9.80 ms | 9.66 ms | 9.81 ms | 9.66 ms |
| | 100mA | 9.62 ms | - | 9.61 ms | 9.85 ms | 9.86 ms | 9.65 ms | 9.78 ms |
| | 1A | 9.60 ms | 9.84 ms | - | 9.65 ms | 9.73 ms | 10.12 ms | 9.91 ms |
| | 10A | 9.76 ms | 9.66 ms | 9.79 ms | - | 9.91 ms | 10.02 ms | 9.63 ms |
| | 50A | 9.60 ms | 9.75 ms | 9.63 ms | 10.02 ms | - | 9.76 ms | 9.81 ms |
| | AUX, 500mV | 9.67 ms | 10.23 ms | 10.36 ms | 9.67 ms | 10.30 ms | - | 10.32 ms |
| | AUX, 5V | 9.97 ms | 9.63 ms | 9.84 ms | 10.47 ms | 10.39 ms | 10.43 ms | - |

- When turning on and off with “:OUTPUT”

| | 1V | 10V | 30V | 100V | 300V | 1000V | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|------------|----------|
| OFF → ON | 32.29 ms | 21.94 ms | 29.29 ms | 21.17 ms | 20.78 ms | 21.40 ms | 21.82 ms | 21.73 ms | 21.06 ms | 27.59 ms | 21.12 ms | 21.06 ms | 21.59 ms |
| ON → OFF | 8.14 ms | 8.24 ms | 8.39 ms | 9.62 ms | 8.55 ms | 8.61 ms | 8.19 ms | 8.30 ms | 8.97 ms | 8.33 ms | 8.21 ms | 8.98 ms | 8.45 ms |

The periods mentioned above are not the period until actual behavior is finished but the one until a communication command is finished.

Actual behavior period is as follows;

OFF → ON: AUX range (approx. 160 ms), Other than the AUX range (approx. 60 ms)

ON → OFF: AUX range (approx. 310 ms), Other than the AUX range (approx. 210 ms)

2-8-2. Communication speed (GPIB)

■ GP-IB

- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:TYPE”

| | | After change | | | | | |
|---------------|--------|--------------|----------|----------|----------|----------|----------|
| | | 1P2W | 1P3W | 3P3W | 3V3A | 3P4W | 1P2WHC |
| Before change | 1P2W | - | 81.23 ms | 85.74 ms | 85.98 ms | 85.50 ms | 85.94 ms |
| | 1P3W | - | - | 95.65 ms | 95.71 ms | 96.76 ms | 96.42 ms |
| | 3P3W | - | 97.02 ms | - | 94.94 ms | 96.35 ms | 96.15 ms |
| | 3V3A | - | 96.35 ms | 95.52 ms | - | 95.27 ms | 92.04 ms |
| | 3P4W | - | 95.44 ms | 96.09 ms | 96.40 ms | - | 96.34 ms |
| | 1P2WHC | - | 96.76 ms | 96.66 ms | 96.88 ms | 96.62 ms | - |

- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:PRIOrity”

| | | After change | |
|---------------|--------|---------------|---------------|
| | | UNIT2 → UNIT3 | UNIT3 → UNIT2 |
| Before change | 1P2W | - | - |
| | 1P3W | 96.35 ms | 91.97 ms |
| | 3P3W | 96.87 ms | 95.17 ms |
| | 3V3A | 93.61 ms | 95.29 ms |
| | 3P4W | 95.94 ms | 94.68 ms |
| | 1P2WHC | 92.87 ms | 93.90 ms |

2-8-3. Communication speed (Ethernet)

■ Ethernet

- When changing the range with “:SOUR:VOLT:RANG”

| | | After change | | | | | |
|---------------|-------|--------------|---------|---------|---------|---------|---------|
| | | 1V | 10V | 30V | 100V | 300V | 1000V |
| Before change | 1V | - | 9.50 ms | 9.51 ms | 9.36 ms | 9.21 ms | 9.52 ms |
| | 10V | 9.41 ms | - | 9.46 ms | 9.52 ms | 9.82 ms | 9.30 ms |
| | 30V | 9.29 ms | 9.42 ms | - | 9.33 ms | 9.59 ms | 9.13 ms |
| | 100V | 9.97 ms | 9.16 ms | 9.32 ms | - | 9.31 ms | 9.33 ms |
| | 300V | 9.35 ms | 9.33 ms | 9.33 ms | 9.42 ms | - | 9.37 ms |
| | 1000V | 9.34 ms | 9.35 ms | 9.50 ms | 9.27 ms | 9.39 ms | - |

- When changing the range with “:SOUR:CURRE:RANG”

| | | After change | | | | | | |
|---------------|------------|--------------|----------|----------|----------|----------|------------|----------|
| | | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
| Before change | 30mA | - | 9.72 ms | 9.86 ms | 9.80 ms | 9.66 ms | 9.81 ms | 9.66 ms |
| | 100mA | 9.62 ms | - | 9.61 ms | 9.85 ms | 9.86 ms | 9.65 ms | 9.78 ms |
| | 1A | 9.60 ms | 9.84 ms | - | 9.65 ms | 9.73 ms | 10.12 ms | 9.91 ms |
| | 10A | 9.76 ms | 9.66 ms | 9.79 ms | - | 9.91 ms | 10.02 ms | 9.63 ms |
| | 50A | 9.60 ms | 9.75 ms | 9.63 ms | 10.02 ms | - | 9.76 ms | 9.81 ms |
| | AUX, 500mV | 9.67 ms | 10.23 ms | 10.36 ms | 9.67 ms | 10.30 ms | - | 10.32 ms |
| | AUX, 5V | 9.97 ms | 9.63 ms | 9.84 ms | 10.47 ms | 10.39 ms | 10.43 ms | - |

- When turning on and off with “:OUTPUT”

| | 1V | 10V | 30V | 100V | 300V | 1000V | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|------------|----------|
| OFF → ON | 32.29 ms | 21.94 ms | 29.29 ms | 21.17 ms | 20.78 ms | 21.40 ms | 21.82 ms | 21.73 ms | 21.06 ms | 27.59 ms | 21.12 ms | 21.06 ms | 21.59 ms |
| ON → OFF | 8.14 ms | 8.24 ms | 8.39 ms | 9.62 ms | 8.55 ms | 8.61 ms | 8.19 ms | 8.30 ms | 8.97 ms | 8.33 ms | 8.21 ms | 8.98 ms | 8.45 ms |

- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:TYPE”

| | | After change | | | | | |
|---------------|--------|--------------|----------|----------|----------|----------|----------|
| | | 1P2W | 1P3W | 3P3W | 3V3A | 3P4W | 1P2WHC |
| Before change | 1P2W | - | 80.00 ms | 91.67 ms | 92.79 ms | 92.57 ms | 92.03 ms |
| | 1P3W | - | - | 98.11 ms | 98.95 ms | 98.65 ms | 98.15 ms |
| | 3P3W | - | 97.90 ms | - | 99.05 ms | 98.80 ms | 97.85 ms |
| | 3V3A | - | 96.99 ms | 96.82 ms | - | 97.38 ms | 96.74 ms |
| | 3P4W | - | 96.35 ms | 96.70 ms | 97.28 ms | - | 96.56 ms |
| | 1P2WHC | - | 98.23 ms | 98.03 ms | 98.86 ms | 96.89 ms | - |

- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:PRiority”

| | | After change | |
|---------------|--------|---------------|---------------|
| | | UNIT2 → UNIT3 | UNIT3 → UNIT2 |
| Before change | 1P2W | - | - |
| | 1P3W | 97.93 ms | 96.71 ms |
| | 3P3W | 97.52 ms | 96.55 ms |
| | 3V3A | 97.13 ms | 96.56 ms |
| | 3P4W | 97.43 ms | 96.56 ms |
| | 1P2WHC | 97.63 ms | 96.74 ms |

2-8-4. Communication speed (USB)

■ Ethernet

- When changing the range with “:SOUR:VOLT:RANG”

| | | After change | | | | | |
|---------------|-------|--------------|---------|---------|---------|---------|---------|
| | | 1V | 10V | 30V | 100V | 300V | 1000V |
| Before change | 1V | - | 7.80 ms | 7.82 ms | 7.31 ms | 7.28 ms | 7.24 ms |
| | 10V | 7.19 ms | - | 7.29 ms | 7.19 ms | 7.25 ms | 7.19 ms |
| | 30V | 7.29 ms | 7.25 ms | - | 7.26 ms | 7.23 ms | 7.30 ms |
| | 100V | 7.26 ms | 7.16 ms | 7.89 ms | - | 7.31 ms | 7.23 ms |
| | 300V | 7.71 ms | 7.25 ms | 7.85 ms | 7.30 ms | - | 7.18 ms |
| | 1000V | 7.25 ms | 7.24 ms | 7.53 ms | 7.32 ms | 7.25 ms | - |

- When changing the range with “:SOUR:CURRE:RANG”

| | | After change | | | | | | |
|---------------|------------|--------------|---------|---------|---------|---------|------------|---------|
| | | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
| Before change | 30mA | - | 7.63 ms | 8.11 ms | 7.48 ms | 7.52 ms | 7.65 ms | 7.60 ms |
| | 100mA | 7.57 ms | - | 7.62 ms | 7.75 ms | 7.61 ms | 7.59 ms | 7.76 ms |
| | 1A | 7.51 ms | 7.53 ms | - | 7.60 ms | 8.05 ms | 7.64 ms | 7.59 ms |
| | 10A | 7.59 ms | 7.59 ms | 7.87 ms | - | 7.60 ms | 7.61 ms | 7.61 ms |
| | 50A | 7.53 ms | 7.61 ms | 7.52 ms | 7.60 ms | - | 7.55 ms | 7.71 ms |
| | AUX, 500mV | 7.69 ms | 7.67 ms | 7.57 ms | 7.62 ms | 7.96 ms | - | 7.67 ms |
| | AUX, 5V | 7.56 ms | 7.54 ms | 7.59 ms | 7.57 ms | 7.58 ms | 7.78 ms | - |

- When turning on and off with “:OUTPUT”

| | 1V | 10V | 30V | 100V | 300V | 1000V | 30mA | 100mA | 1A | 10A | 50A | AUX, 500mV | AUX, 5V |
|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|----------|------------|----------|
| OFF → ON | 22.15 ms | 22.11 ms | 23.41 ms | 24.27 ms | 23.80 ms | 22.06 ms | 22.15 ms | 22.30 ms | 22.22 ms | 22.19 ms | 23.93 ms | 22.57 ms | 23.95 ms |
| ON → OFF | 6.37 ms | 6.32 ms | 6.74 ms | 6.39 ms | 6.16 ms | 6.47 ms | 6.26 ms | 6.22 ms | 6.30 ms | 6.19 ms | 6.20 ms | 6.25 ms | 6.23 ms |

- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:TYPE”

| | | After change | | | | | |
|---------------|--------|--------------|-----------|-----------|-----------|-----------|-----------|
| | | 1P2W | 1P3W | 3P3W | 3V3A | 3P4W | 1P2WHC |
| Before change | 1P2W | - | 596.15 ms | 596.10 ms | 596.11 ms | 595.94 ms | 596.28 ms |
| | 1P3W | - | - | 599.83 ms | 599.87 ms | 599.85 ms | 599.79 ms |
| | 3P3W | - | 599.82 ms | - | 599.79 ms | 601.83 ms | 599.83 ms |
| | 3V3A | - | 605.80 ms | 601.79 ms | - | 599.82 ms | 599.85 ms |
| | 3P4W | - | 599.83 ms | 599.85 ms | 599.85 ms | - | 603.85 ms |
| | 1P2WHC | - | 599.80 ms | 599.78 ms | 599.83 ms | 599.80 ms | - |


- Time period until the ANS of the status register becomes 0 when changing the device structure with “:WIRing:PRIOrity”

| | | After change | |
|---------------|--------|---------------|---------------|
| | | UNIT2 → UNIT3 | UNIT3 → UNIT2 |
| Before change | 1P2W | - | - |
| | 1P3W | 599.82 ms | 598.91 ms |
| | 3P3W | 599.88 ms | 599.16 ms |
| | 3V3A | 599.85 ms | 598.92 ms |
| | 3P4W | 599.90 ms | 599.16 ms |
| | 1P2WHC | 599.87 ms | 599.11 ms |

3-1. Power error at three-phase power calibration

Overview

In the case of three-phase three wire system, there is a phase difference of 30° between voltage and current as indicated in the right diagram, and the characteristic of power error is different from other wiring system. This is the value in a representative range.



Phase diagram
3P3W

Condition settings

LS3300 3P3W

Range = 300 V, 10 A

Output setting value = $100\sqrt{3}$ V = 173.205 V
(voltage between wires)

10 A (wire current)

Frequency = 50/60 Hz

Three phase apparent power

Three phase apparent power = $\sqrt{3} \times \text{voltage between wires} \times \text{wire current}$ VA

View about error

Error against each LS3300 = 0.04% of VA + 0.005% of Range + PWRerror% of VA


* PWRerror is the computational expression of addition value to active power

accuracy by phase error ($\Delta\Phi$)

$\text{PWRerror} (\%) = 100 \times \{\cos\Phi - \cos(\Phi + \Delta\Phi)\}$

Whole error is the total of each LS3300's error.

Computational expression of power error of 3P3W



3-2. Stability of large 1P2W current

■ Overview

This is the fluctuation of the output error of a representative range when three LS3300s performs synchronous output with the wiring setting of "Hi Current" connecting the power meter WT3000E. The period of the plotted data is one hour from one minute after the output is turned on.

■ Used device

Power Meter WT3000E, CT1000

■ Device settings


WT3000E Update rate = 500 ms

LS3300 Range = 150 A

Output setting value = 100% of the range

Frequency = 60 Hz

■ Measurement result



3-3-1. Example of calibration using AUX terminal

■ Overview

This is the calibration example of the power quality analyzer YOKOGAWA CW500 when the AUX terminal of the LS3300 is used.

■ Detailed explanation

Because the YOKOGAWA CW500 uses a current clamp for current input, the input of the current clamp is connected to the AUX terminal of the LS3300.

The LS3300 outputs voltage as usual, outputs voltage converted at the conversion ratio of a current clamp from the AUX terminal and performs calibration in the same manner as the calibration of a power meter.

■ Used device

YOKOGAWA CW500 Power quality analyzer

■ Device settings

LS3300 Range = 100 V, AUX 500 mV

Output setting value = Arbitrary setting value of a current clamp to be used

Frequency = 50 Hz


CW500 Range = 600 V, Clamp 1 A/1 mV: 96064 (500 A)

■ Actual measurement screen

<Connection example of LS3300 and CW500>



<② Conversion cable for current clamp>



3-3-2. Notes on connection of AUX terminals

■ Overview

This subsection shows an error cause when AUX is connected.

■ Detailed explanation

The common terminals of the current input terminal and the external current sensor input terminal are connected internally in the YOKOGAWA WT series. If the current terminal of the LS3300 and the AUX terminal are connected simultaneously, an error develops in the AUX output.

When you calibrate a power meter whose current input and external current sensor input are connected internally, if you connecting the AUX terminal, please remove the connection of the current terminal.

■ Used device

Power meter YOKOGAWA WT3000

■ Device settings

WT3000 Range = EXT 500 mV, EXT 5 V

Update rate = 1 S

LS3300 Range = AUX 500 mV, AUX 5 V

Output setting value = 125% of each range

Frequency = 60 Hz

■ Measurement result


| LS3300 range | Setting value | Only AUX is connected | Current and AUX are connected | Error (ppm) |
|--------------|---------------|-----------------------|-------------------------------|-------------|
| AUX 500 mV | 625.00 mV | 624.984 mV | 625.143 mV | 317 |
| AUX 5 V | 6.2500V | 6.25017 V | 6.25032 V | 30 |

***Voltage and current output are the same.**

3-4. Change of wiring system via communication and setting procedure

■ Overview

The procedure of setting change of wiring system/priority unit using a communication command is as follows:



4-1. Safety of output terminal

- **Overview**

This section shows the safety of the output terminals.

- **Voltage terminal**

Even though the voltage output terminals are safety terminals, it generates high voltage, maximum 1250 Vrms. If you operate it in a wrong manner, there is a possibility that a serious accident affecting people's lives may occur. To prevent electric shock, be sure to read the user's manual before use.

- **Current terminal**

Because the compliance voltage is 22 Vrms or less when the LS3300 generates current, and the voltage is lower than the danger voltage 46.7Vpeak, so the current output terminals are safety. In case that the Lo terminal of the current is not connect into the earth (Lo to Earth = off), because 12 Vpeak is the maximum additional voltage against the earth and the voltage between Hi terminal and the case is under the danger voltage 46.7 Vpeak, the current output terminals are safety.

4-2. Example of error cause during voltage/current generation (1)

- **Overview**

This is the influence of the interference when the LS3300 outputs the combination of high voltage and small current or high current and low voltage at simultaneous output of voltage and current.

- **Used device**

Power Meter WT3000E

- **Device settings**

WT3000E Update rate = 500 ms



- **Notes on the output of high voltage and large current**

When high voltage and small current are output together, a current output value sometimes fluctuates according to output frequency and phase due to the influence of the electric field of high voltage output. To reduce the influence of the electric field, please twist voltage and current cables respectively and construct routing avoiding the influence of the electric field.



When large current and low voltage are output together, a voltage output value sometimes fluctuates according to output frequency and phase due to the influence of the magnetic field of large current output. To reduce the influence of the magnetic field, please twist voltage and current cables respectively and construct routing avoiding the influence of the magnetic field.

4-2. Example of error cause during voltage/current generation (2)

- Variation of current output by high voltage output
 - Variation of 30 mA when 1250 V is output*

| Frequency | When voltage cables and current cables are apart from each other | When voltage cables and current cables are close together |
|-----------|---|---|
| 60 Hz | No impact | 100 ppm |
| 400 Hz | 90 ppm | 620 ppm |
| 1200Hz | 250 ppm | 1800 ppm |
| |  <p>More than 10 cm</p> |  |

- Variation of voltage output by large current output
 - Variation of 1 V when 62.5 A is output*

| Frequency | When voltage cables and current cables are apart from each other | When voltage cables and current cables are close together |
|-----------|--|--|
| 60 Hz | No impact | 50 ppm |
| 400 Hz | 70 ppm | 400 ppm |
| 1200 Hz | 260 ppm | 1300 ppm |
| |  <p>More than 10 cm</p> |  |

*The cables are twisted.

4-2. Example of error cause during voltage/current generation (3)

■ Overview

This is the error cause during voltage generation.


■ Detailed explanation

Because the LS3300 is a generator with two terminal connection, an error of output value is caused by output current.

The guaranteed output resistance of a terminal part (including the contact resistance of a safety terminal in a recommended cable) is 6 mΩ (except for the 100 mV range). The voltage drop of 6 mV accordingly occurs at the terminal part when output current is 1 A. (diagram [1]).

In addition, a voltage drop caused by a cable to a DUT also makes an error happen (diagram [2]).

You therefor have to monitor voltage at the DUT end if output current exerts an influence and adjust output settings to make the LS3300 output voltage you want (diagram [3]).



4-3. Maximum capacity and inductive load

■ Overview

This is the maximum capacity and the inductive load of each range of the LS3300A.

■ Overview

The following tables show the range where oscillation doesn't occur on the premise that output is within the compliance.

If the LS3300 performs output while a large inductive load or a load of resonance circuit is connected to the current terminal, it is possible that oscillation occurs and large voltage is output to internal output circuit. If oscillation occurs, turn off output ASAP.

Because there is a possibility that oscillation occurs, do not connect an inductive load at voltage output and a capacity load at current output.

| Range | Max. capacity load | Range | Max. inductive load |
|--------|--------------------|------------|---------------------|
| 1 V | 10 μ F | 30 mA | 1 mH |
| 10 V | 10 μ F | 100 mA | 1 mH |
| 30 V | 1 μ F | 1 A | 1 mH |
| 100 V | 1 μ F | 10 A | 1 mH |
| 300 V | 0.1 μ F | 50 A | 1 mH |
| 1000 V | 0.01 μ F | Range | Max. capacity load |
| | | AUX 500 mV | 1000 pF |
| | | AUX 5 V | 1000 pF |

4-4. Power meter in which voltage and current terminals are connected internally

■ Overview


This is the connection to a power meter where voltage and current terminals are connected internally.

■ Detailed explanation

When voltage and current terminals are connected internally in a 3P3W power meter, the LS3300 has to input current to the power meter isolating with a CT because the maximum application voltage between the Lo terminal of current output and the ground is 12 Vpk.

We recommend to use another power meter for reference because an error may occur due to the current transformation ratio and the phase of the CT.

■ Example of calibration



4-5. Example of calibration of clamp-on power meter

■ Overview

This section shows how to set the LS3300 for the calibration of a clamp-on power meter.

With the use of a coil, the LS3300 can perform a test equivalent for energization of output current x turn count.

■ Used device

Clamp-on Power Checker YOKOGAWA CW10

Test coil for a clamp-on meter (wire material: PEWΦ2.3 mm, the number of turns: 60)

■ Device settings

LS3300 Voltage terminal: No load (direct connection to the voltage input terminal of the CW10)

Current terminal: Connection to a test coil for a clamp-on meter

Voltage range = 1000 V, Voltage output setting value = 100%


Current range = 10 A, current output setting value = 100%

(Current at the center of the coil: 10 A × 60 turns = 600 A)

Wiring system = 1P2W

Reference signal = INT, Frequency = 60 Hz

■ Example of calibration



*Display value of clamp-on power meter: 600.0 kW

Voltage between LS3300's current terminals during output mentioned above: 2.1 V (maximum output spec: approx. 3V)

4-6. Indication of output value


■ Overview

This section explains differences from the 2558A in the indication of the output values of LS3300's voltage, current, power and apparent power.

■ Differences between the 2558A and the LS3300

| | LS3300 | 2558A |
|---------------------|--|---|
| Indication value | Output value = calculated value of output level x level ratio While output varies, "STABILIZEING" is displayed at the upper right of the screen and it disappears when the output becomes stable. | Measurement value: When the OUTPUT switch is turned on, the value fluctuates until it becomes stable. Indication sometimes changes according to the state of a load during output. |
| Indication location | The field of Output Value on the screen as shown in the diagram 1 below | OUTPUT indication in the right side of the front panel as shown in the diagram 2 below |
| Others | When "*" is displayed next to the indication of output value, the target signal is not being output. | When the OUTPUT switch is turned off, "— — — — —" is displayed. |

■ Indication example of the LS3300: Diagram 1



Indication example of the 2558A: Diagram 2



4-7. Sweep behavior

■ Overview

This is the comparison of the sweep function of the 2558A and the LS3300.

| Item | 2558A | LS3300 |
|--------------------|------------------------------------|--|
| Sweep target | Voltage/current/frequency | Voltage/current |
| Sweep range | 0 to 120% of main setting value | Selectable from 0 to 100%/105%/110%/120% of output level |
| Sweep time | 8 s/16 s/32 s/64 s | ← |
| Sweep setting item | Output division setting, deviation | Output level ratio |
| Sweep direction | UP/DOWN | ← |

4-8. LINE synchronization

■ Overview

This section explains the LINE synchronization function.

With this function, the LS3300 can synchronize output frequency with the commercial power supply frequency.

It can be used for checkup of the interference from a power supply.

■ Detailed explanation

When a specified frequency of an AC generator and the frequency of commercial power supply are almost the same, a beat that the indication value of a calibration target increases and decreases at intervals of several tens seconds occurs on rare occasion. There are following two methods to prevent beat with the commercial power supply:

① Changing the setting frequency

② Synchronizing the output of a generator with the commercial power supply

The LINE synchronization achieves the latter method.

When the reference signal of the LS3300 is set at INT or EXT, if suspecting the occurrence of a beat with the commercial power source, you can confirm the problem by changing to the LINE synchronization.

However, the phase difference between the commercial power source and output is not zero.

■ Application examples (external magnetic field test)

With the LINE synchronization function, it is possible to output voltage/current of an arbitrary phase against the commercial power supply.

The JIS standard test confirming the influence of the external magnetic field prescribes that “the frequency is the same as the one that drives the instrument and the phase exerts the maximum influence on the instrument.”

The commercial power source is generally used for a coil generating a magnetic field.

With the LINE synchronization function, you can confirm the influence of the external magnetic field by applying the output whose phase is changed with reference to the commercial power source to a device under test.