User's Manual

Model 751535-J5 RACK MOUNTING KIT For an JIS-compliant Single-housing Rack

1. Checking the Parts in the Kit

Confirm that all parts listed below are included in the kit. If any part is shorted or any abnormality is found in the appearance of the parts, contact our sales representative where you bought the instrument. List of parts included in the kit.

Right side bracket (1 pc), Left side bracket (1 pc), Upper bracket (1 pc), Flat-head screws (M5, 14 mm long, 6 pcs), Flat-head screws (M3, 8 mm long, 2 pcs), sealing labels (8 pcs)

2. Removing the Parts from the Instrument (per each set)

Removing the Handle

Note: There is a chance that the handle cannot be reattached once it has been removed from the instrument.


- 1. Remove the four covers (a) from the handle by firmly inserting a screwdriver in the direction of the four arrows marked A. If the covers stick to the handle, use another tool to widen the opening of each cover horizontally and then retry taking off the cover.
- 2. To remove the handle, unscrew the four screws (b) which fix the handle to the instrument.

Removing the Feet on the Bottom

- 1. When a black pin (i) is included in each foot (c), pry out the four pins with a thin screwdriver as shown in the figure below.
- 2. Press each foot (c) in the direction of arrow B, that is, toward the rear of the instrument, to slide off the four feet (c).
- **3.** Press each foot (c) in the direction of arrow C, that is, downward, to remove the four feet (c). After the front feet (c) are removed, the support hardware (d) is also detached.

Removing the Sealing Labels

- 1. Remove the four sealing labels (e).
- 2. Unplug the two blind plugs (f).



Printed in Japan



3. Joining the Instruments with Mounting Hardware

- 1. With the eight labels (1), seal the four openings on the bottom plate to fix the feet, and seal the four openings on the side plate to fix the handle.
- 2. Using three flat-head screws (M5, 14 mm long) (3), screw the left side bracket (2) to the instrument to be located on the left side of the rack.


Reference torque value: 2.8 to 2.9 N·m

3. Using three other flat-head screws (M5, 14 mm long) (3), screw the right side bracket (4) to the right side of the instrument to be located on the right side of the rack.

Reference torque value: 2.8 to 2.9 N·m

4. Screw the upper bracket (5) to the left and right side brackets with two flat-head screws (6) (M3, 8 mm long).

Reference torque value: 0.6 to 0.7 N·m




^{*1} This hole is not used when rack frame is attached to a WT1801E, WT1802E, WT1803E, WT1804E, WT1805E, or WT1806E model with the /PD option.

4. Mounting the Joined Instruments to the Rack

Dimensions after mounting to the rack are illustrated below.

Note: 1. Be sure to place support beneath the rack where the support will not cover the vent openings on the bottom of the instruments.

2. Provide a clearance of at least 80 mm around the instrument for ventilation.



2/2 IM 751535-J5-01E

^{*2} This hole is not used when rack frame is attached to models other than those listed above.