

Name of Accreditation Program	JCSS Accreditation Program
Accreditation Identification	JCSS 0372 Calibration
Name of Conformity Assessment Body	Yokogawa Test & Measurement Corporation Common Technology & QA Division Product Compliance Dept. Calibration Sec.
Name of Legal Entity	Yokogawa Test & Measurement Corporation JCN 5012801006673
Inquiry Point	Common Technology & QA Division Product Compliance Dept. Calibration Sec. TEL: +81-42-690-8809      FAX: +81-42-690-8827

\*JCN: Japan Corporate Number



International Accreditation Japan

## Information on Accredited Calibration Laboratory

Date of the update of the information : 2026-04-28

Accreditation Identification: JCSS 0372 Calibration

Name of Calibration Laboratory : Yokogawa Test & Measurement Corporation  
Common Technology & QA Division  
Product Compliance Dept. Calibration Sec.

Location of Calibration Laboratory : 155 Takamuro-cho, Kofu-shi, Yamanashi  
400-8558, JAPAN

Name of Legal Entity: Yokogawa Test & Measurement Corporation

Conformance Accreditation Standard: ISO/IEC 17025:2017

Expiry Date of Accreditation : 2028-03-05

Effective Date of Accreditation: 2024-03-06

General Field of Calibration: Time & Frequency & Rotational speedDate of Initial Accreditation of the Field: 2024-03-06Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range	Expanded Uncertainty (Level of Confidence Approximately 95 %)
Time & Frequency Counter, etc.	Frequency Counter	200 kHz	$1.5 \times 10^{-6}$

#All Calibration Procedures are in-house procedures developed by this laboratory.

General Field of Calibration: Electricity (Direct Current & Low Frequency)Date of Initial Accreditation of the Field: 2024-03-06Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated		Range		Expanded Uncertainty (Level of Confidence Approximately 95 %)
Direct Current & Low Frequency Measuring Equipment, etc.	DC Resistor	From 1 $\Omega$ up to 400 $\Omega$		25 ppm + 4 m $\Omega$
	DC Voltage Source	From 10 mV up to 1000 V From -1000 V up to -10 mV		25 ppm or 1.7 $\mu$ V Which is larger
	DC Voltage Measuring Equipment	From 1 mV up to 42 V From -42 V up to -1 mV		0.02 % + 1 $\mu$ V
		More than 42 V up to 1000 V From -1000 V less than -42 V		0.08 % or 16 $\mu$ V Which is larger
	Direct Current Source	100 $\mu$ A, -100 $\mu$ A		8.0 nA
		1 mA, -1 mA		80 nA
		10 mA, -10 mA		0.80 $\mu$ A
		30 mA, -30 mA		3.0 $\mu$ A
		100 mA, -100 mA		7.0 $\mu$ A
		1 A, -1 A		0.20 mA
		10 A, -10 A		4.0 mA
		30 A		10 mA
	AC Voltage Source	1 kHz 400 Hz 60 Hz	100 mV, 1 V, 10 V, 50 V, 100 V, 300 V, 1000 V	0.015 %
	Alternating Current Source	60 Hz	100 mA	0.022 mA
			1 A	0.23 mA
5 A			1.5 mA	
10 A			2.3 mA	
20 A			4.6 mA	
		50 A	12 mA	

Effective Date of Accreditation: 2024-03-06

Direct Current & Low Frequency Measuring Equipment, etc.	Temperature Indicator calibration equipment	Resistance thermometer Sensor	390.48 $\Omega$ (850 °C)	25 ppm + 4 m $\Omega$	
			100 $\Omega$ (0 °C)		
			18.52 $\Omega$ (-200 °C)		
		Temperature Indicator	Thermocouple K, without Reference Junction	52.41 mV (1300 °C)	1.7 $\mu$ V
				-6.404 mV (-250 °C)	
		Temperature Indicator	Thermocouple K without Reference Junction	-5.891 mV, 52.410 mV (-200 °C, 1300 °C)	10 $\mu$ V
			Thermocouple E without Reference Junction	-8.825 mV, 61.017 mV (-200 °C, 800 °C)	
			Thermocouple T without Reference Junction	-5.603 mV, 20.872 mV (-200 °C, 400 °C)	

#All Calibration Procedures are in-house procedures developed by this laboratory.