

Please make the following alterations to the User's Manual IM WT1801-01EN on the included CD-ROM.

■ **7-5 page "Notes about the Numeric Data Display"**

Please change the followings

-
-
- When the Crest Factor Is Set to CF3
When Urms, Uac, Irms, or Iac is 0.3% or less When Umn, Urmn, Imn, or Irmn is 2% or less
-
-

■ **16-3 page "Wiring System during High Speed Data Capturing"**

Please change the followings

-
- 1P2W: Single-phase, two-wire system (DC signal)
- 3P4W: Three-phase, four-wire system
- 3P3W (3V3A): Three-voltage, three-current method